

DWYER INSTRUMENTS, INC.

102 INDIANA HIGHWAY 212 | P.O. BOX 373 MICHIGAN CITY, IN 46360-1956

DWYER-INST.COM

EU - DECLARATION OF CONFORMITY 04

We.

Dwyer Instruments, Inc. 102 Indiana Highway 212 Michigan City, IN 46360 USA +1-219-879-8868

declare under our sole responsibility that our Series 626, 627, and 628 Industrial Pressure Transmitter to which this declaration relates is in conformity with the following Directives and harmonized Standards:

Directive 2011/65/EU Restriction of Certain Hazardous Substances in Electrical and Electronic Equipment + Directive (EU) 2015/863 Amendment to Annex II

Directive 2014/30/EU Electromagnetic Compatibility (EMC)

EN 61326-1:2013 Electrical Equipment for Measurement, Control and Laboratory Use - EMC Requirements - Part 1: General Requirements

IEC 61000-4-2:2008 Electromagnetic Compatibility (EMC) - Part 4-2: Testing and Measurement Techniques - Electrostatic Discharge Immunity Test

IEC 61000-4-3:2006+A1:2008+A2:2010 Electromagnetic Compatibility (EMC) - Part 4-3: Testing and Measurement Techniques - Radiated, Radio-Frequency, Electromagnetic Field Immunity Test

IEC 61000-4-4:2012 Electromagnetic Compatibility (EMC) - Part 4-4: Testing and Measurement Techniques - Electrical Fast Transient/Burst Immunity Test

IEC 61000-4-5:2005 Electromagnetic Compatibility (EMC) - Part 4-5: Testing and Measurement Techniques - Surge Immunity Test

IEC 61000-4-6:2013 Electromagnetic Compatibility (EMC) - Part 4-6: Testing and Measurement Techniques - Immunity to Conducted Disturbances, Induced By Radio-Frequency Fields

EN 55011:2016 Industrial, Scientific and Medical Equipment. Radio-Frequency Disturbance Characteristics. Limits and Methods of Measurement

The authorized representative located within the Community is:

Dwyer Instruments Ltd Unit 16 The Wye Estate, London Road High Wycombe, HP11 1LH-U.K. +44 (0) 1494 461707

On behalf of Dwyer Instruments, Inc.,

Doug McCall Senior Regulatory Engineer February 5, 2019 Dwyer Instruments, Inc. 102 Indiana Highway 212 Michigan City, IN 46360, USA +1-219-879-8868