

CE



04

DECLARATION OF CONFORMITY

We,

Dwyer Instruments, Inc.
P.O. Box 373
Michigan City, IN 46361, USA
(219) 897-8868

declare under our sole responsibility that our Digihelic Models DH-001 through DH-017 and Digihelic II Models DHII-001 through DHII-017 to which this declaration relates are in Conformity with the following standards:

- EN 61000-4-2 (2001) Electromagnetic Compatibility (EMC) – Part 4-2: Testing Measurement Techniques – Electrostatic Discharge Immunity Test
- EN 61000-4-3 (2003) Electromagnetic Compatibility (EMC) – Part 4-3: Testing Measurement Techniques – Radio Frequency, Electromagnetic Field Immunity Test
- EN 61000-4-4 (1995) Electromagnetic Compatibility (EMC) – Part 4-4: Testing Measurement Techniques – Electrical Fast Transient/Burst Immunity Test
- EN 61000-4-5 (2001) Electromagnetic Compatibility (EMC) – Part 4-4: Testing Measurement Techniques – Surge Immunity Test
- EN 61000-4-6 (2003) Electromagnetic Compatibility (EMC) – Part 4-6: Testing Measurement Techniques – Immunity To Conducted Disturbances, Induced By Frequency Fields
- EN 61000-4-11(2004) Electromagnetic Compatibility (EMC) – Part 4-11: Testing Measurement Techniques – Voltage Dips, Short Interruptions, and Voltage Variations Test.
- EN 61010 –1:90+A1:92+A2:95 Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use.

Per the provisions of the Electromagnetic Compatibility Directive 89/336/EEC and the Low Voltage Directive 73/23/EEC

The authorized representative located within the Community is:

Dwyer Instruments Ltd
Unit 16, The Wye Estate, London Road
High Wycombe, Bucks HP11 1LH-U.K.
(+44) (0) 1494 461707

On behalf of Dwyer Instruments, Inc.

A handwritten signature in black ink, appearing to read "Larry Roth".

Senior Project Engineer

Michigan City, Indiana, USA
11/05/04