



04

## EC - DECLARATION OF CONFORMITY

We,

Dwyer Instruments, Inc.  
P.O. Box 373  
Michigan City, IN 46361, USA  
(219) 879-8868

declare under our sole responsibility that our MP Series Differential Pressure Switch/Gage to which this declaration relates are in conformity with the following EC Directives and harmonized standards:

### Directive 89/336/EEC (EMC)

- EN 61000-4-2 (2001) Electromagnetic Compatibility (EMC) – Part 4-2: Testing Measurement Techniques – Electrostatic Discharge Immunity Test
- EN 61000-4-3 (2002) Electromagnetic Compatibility (EMC) – Part 4-3: Testing Measurement Techniques – Radiated, Radio Frequency, Electromagnetic Field Immunity Test
- EN 61000-4-4 (1995) Electromagnetic Compatibility (EMC) – Part 4-4: Testing Measurement Techniques – Electrical Fast Transient/Burst Immunity Test
- EN 61000-4-5 (2001) Electromagnetic Compatibility (EMC) – Part 4-4: Testing Measurement Techniques – Surge Immunity Test
- EN 61000-4-6 (2003) Electromagnetic Compatibility (EMC) – Part 4-6: Testing Measurement Techniques – Immunity To Conducted Disturbances, Induced By Frequency Fields
- EN 61326 (2002) Electrical Equipment For Measurement, Control And Laboratory Use EMC Requirements
- EN 55011 (1998) Industrial, Scientific And Medical (ISM) Radio Frequency Equipment – Electromagnetic Disturbance Characteristics – Limits And Methods Of Measurement

### Directive 72/23/EEC (Low Voltage)

- EN 61010-1 (1995) Safety Requirements For Electrical Equipment For Measurement, Control And Laboratory Use

The authorized representative located within the Community is:

Dwyer Instruments Ltd  
Unit 16, The Wye Estate, London Road  
High Wycombe, Bucks HP11 1LH-U.K.  
(+44) (0) 1494 461707

On behalf of Dwyer Instruments, Inc.

Senior Project Engineer

Michigan City, Indiana, USA  
5/28/04