

# EU - DECLARATION OF CONFORMITY **13**

We,

Dwyer Instruments, Inc.  
102 Indiana Highway 212  
Michigan City, IN 46360 USA  
+1-219-879-8868

declare under our sole responsibility that our Series 607 Differential Pressure Transmitter to which this declaration relates is in conformity with the following Directives and harmonized Standards:

Directive 2011/65/EU Restriction of Certain Hazardous Substances in Electrical and Electronic Equipment

Directive 2014/30/EU Electromagnetic Compatibility (EMC)

IEC 61326-1:2013 Electrical Equipment for Measurement, Control and Laboratory Use - EMC Requirements - Part 1: General Requirements

IEC 61236-2-3:2012 Electrical Equipment for Measurement, Control and Laboratory Use - EMC Requirements - Part 2-3: Particular Requirements - Test Configuration, Operational Conditions and Performance Criteria for Transducers with Integrated or Remote Signal Conditioning.

IEC 61000-4-2:2008 Electromagnetic Compatibility (EMC) - Part 4-2: Testing and Measurement Techniques - Electrostatic Discharge Immunity Test

IEC 61000-4-3:2010 Electromagnetic Compatibility (EMC) - Part 4-3: Testing and Measurement Techniques - Radiated, Radio-Frequency, Electromagnetic Field Immunity Test

IEC 61000-4-4:2012 Electromagnetic Compatibility (EMC) - Part 4-4: Testing and Measurement Techniques - Electrical Fast Transient/Burst Immunity Test

IEC 61000-4-5:2014 Electromagnetic Compatibility (EMC) - Part 4-5: Testing and Measurement Techniques - Surge Immunity Test

IEC 61000-4-6:2013 Electromagnetic Compatibility (EMC) - Part 4-6: Testing and Measurement Techniques - Immunity to Conducted Disturbances, Induced By Radio-Frequency Fields

The authorized representative located within the Community is:

Dwyer Instruments Ltd  
Unit 16 The Wye Estate, London Road  
High Wycombe, HP11 1LH-U.K.  
+44 (0) 1494 461707

On behalf of Dwyer Instruments, Inc.,



Doug McCall  
Senior Regulatory Engineer  
September 26, 2018

Dwyer Instruments, Inc.  
102 Indiana Highway 212  
Michigan City, IN 46360, USA  
+1-219-879-8868