





102 Indiana Highway 212 | P.O. Box 373 | Michigan City, IN 46360

dwyer-inst.com

## **EU - DECLARATION OF CONFORMITY**

We,

Dwyer Instruments, Inc. 102 Indiana Highway 212 Michigan City, IN 46360 USA +1-219-879-8868

declare under our sole responsibility that our Series LDPM Digital Differential Pressure Manometer to which this declaration relates is in conformity with the following Directives and harmonized standards:

Directive 2011/65/EU Restriction of Certain Hazardous Substances in Electrical and Electronic Equipment

Directive 2014/30/EU Electromagnetic Compatibility (EMC)

EN 61326-1:2013 Electrical Equipment for Measurement, Control and Laboratory Use - EMC Requirements

EN 61000-3-2: 2014 Electrical Equipment for Measurement (EMC) - Part 3-2: Limits - Limits for harmonic current emissions

EN 61000-3-3: 2013 Electrical Equipment for Measurement (EMC) - Part 3-3: Limits - Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems

EN 61000-4-2:2009 Electromagnetic Compatibility (EMC) - Part 4-2: Testing and Measurement Techniques - Electrostatic Discharge Immunity Test

EN 61000-4-3:2006+A1:2008+A2:2010 Electromagnetic Compatibility (EMC) - Part 4-3: Testing and Measurement Techniques -Radiated, Radio-Frequency, Electromagnetic Field Immunity Test

EN 61000-4-4:2012 Electromagnetic Compatibility (EMC) - Part 4-4: Testing and Measurement Techniques - Electrical Fast Transient/Burst Immunity Test

EN 61000-4-5:2014 Electromagnetic Compatibility (EMC) - Part 4-5: Testing and Measurement Techniques - Surge Immunity Test

EN 61000-4-6:2014 Electromagnetic Compatibility (EMC) - Part 4-6: Testing and Measurement Techniques - Immunity to Conducted Disturbances, Induced By Radio-Frequency Fields

EN 61000-4-8:2010 Electromagnetic Compatibility (EMC) - Part 4-8: Testing and Measurement Techniques - Power Frequency Magnetic Field Immunity Test

EN 61000-4-11:2004 Electromagnetic Compatibility (EMC) - Part 4-11: Testing and measurement techniques voltage dips, short interruptions and voltage variations immunity tests.

The authorized representative located within the Community is:

**Dwyer Instruments Ltd** Unit 16 The Wye Estate, London Road High Wycombe, HP11 1LH-U.K. +44 (0) 1494 461707

On behalf of Dwyer Instruments, Inc.,

Jennifer L. Williams Associate Regulatory Engineer 11/28/2016

Dwver Instruments, Inc. 102 Indiana Highway 212 Michigan City, IN 46360, USA +1-219-879-8868

DII\_EC\_CEDoC\_v7 Page 1 of 1