



Member of the FM Global Group

FM Approvals
1151 Boston-Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781 762 9375 www.fmapprovals.com

CERTIFICATE OF COMPLIANCE

HAZARDOUS LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

475-a. Manometer.

IS / I / 1 / ABCD/T4 Ta = 70°C

a = Pressure ranges 00, 0, 1, 2, 3, 4, 5, 6, 7.

Special Conditions of Use:

1. For use with Duracell battery P/N MN1604, Eveready P/N 522 or Energizer battery P/N EN22.

477-a. Manometer.

IS / I / 1 / ABCD/T4 Ta = 70°C

a = Pressure ranges 1, 2, 3, 4, 5, 6, 7.

Special Conditions of Use:

1. For use with Duracell battery P/N MN1604, Eveready P/N 522 or Energizer battery P/N EN22.

Equipment Ratings:

Intrinsically Safe for use in Class I, Division 1, Groups A, B, C and D; Temperature Class T4 Ta = 70°C;
indoor Hazardous (Classified) Locations .

FM Approved for:

Dwyer Instruments Inc.
102 Indiana Hwy 212, Box 373, Michigan City IN 46361

This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

Class 3600	1998
Class 3610	1999
Class 3810	2005

Original Project ID: 1D9A7.AX

Approval Granted: August 14, 1998

Subsequent Revision Reports / Date Approval Amended

Report Number	Date	Report Number	Date
3025500	<i>Sept. 20, 2006</i>		

FM Approvals LLC



Robert L. Martell, Jr.
Assistant Vice President

Sept. 20, 2006

Date