



GIG



[1] **EU-TYPE EXAMINATION CERTIFICATE**

[2] Equipment and protective systems intended for use in potentially explosive atmospheres. Directive 2014/34/EU

[3] EU – type examination certificate (module B):

**KDB 17ATEX0056X**

**issue 0**

[4] Equipment:

**Smart Pressure Transmitter type 3400-xx-yy,**

**Smart Differential Pressure Transmitter type 3500-xx-yy**

[5] Manufacturer:

**DWYER INSTRUMENTS INC.**

[6] Address:

**102 Indiana Highway 212, Michigan City, IN 46360, USA**

[7] This product and any acceptable variation thereto is specified in the schedule to this certificate.

[8] Główny Instytut Górnictwa, Notified Body number 1453 in accordance with Directive 2014/34/EU of 26 February 2014, certifies that this equipment or protective system has been found to comply with the Essential Health and Safety Requirements relating to the design and construction of equipment and protective systems intended for use in potentially explosive atmospheres given in Annex II to the Directive 2014/34/EU. The examination and test results are recorded in confidential report **KDB Nr 17.072 [T-7483]**

[9] Compliance with the Essential Health and Safety Requirements has been met by compliance with:

**EN 60079-0:2012 + A11:2013; EN 60079-1:2014;**

**EN 60079-11:2012; EN 60079-26:2015; EN 60079-31:2014**

[10] In case if the sign „X“ is placed after the certificate number, it indicates special conditions for safe use, specified in the schedule to this certificate.

[11] This EU-type examination certificate relates only to the construction, evaluation and tests of product accordance with Directive 2014/34/EU. The certificate does not include other requirements of the Directive relating to manufacturing process and putting into the market of the equipment or protective device.

[12] Marking of the equipment shall include:



**II 1/2G Ex ia/db IIC T6/T5 Ga/Gb**



**II 1/2D Ex ia/tb IIIC T85°C/T100°C Da/Db**

or



**II 2G Ex ia/db IIC T6/T5 Gb**



**II 2D Ex ia/tb IIIC T85°C/T100°C Db**



**KDBEX.eu**

mgr inż. Piotr Madej

ATEX Certification Specialist



KIEROWNIK  
Zespołu Certyfikacji Wytrobów  
KD "BARBARA" Mikołów  
dr hab. inż. Krzysztof Cybulski, prof. GIG

Date of issue: **05.09.2017 r.**

Page 1 of 3

Główny Instytut Górnictwa, 40-166 Katowice, Plac Gwarków 1, POLAND, www.gig.eu

(Certification Body-Certification Team-Kopalnia Doświadczalna "Barbara" Mikołów)

Certification Body accredited by PCA, Nr AC038

This certificate may be reproduced only in its entirety with schedule. The next issue of the certificate replaces the earlier editions.

Issue 0 is the initial certification. The document without signatures and seals is invalid.



**[15] Description:**

Pressure transmitters type 3400-xx-yy and differential pressure transmitters type 3500-xx-yy work by converting proportional to the measured pressure resistance changes of piezoresistive bridge, located in the single crystal of silicon diaphragm, into a standard current signal 4-20 mA with HART communications signal.

The basic units of the transmitter is a measuring head (Ex i) with a silicon diaphragm sensor. Measuring head can be equipped with different pressure connections. Inside the head there is the "pressure chamber" filled with manometer liquid. On the side of measured medium it is limited by a diaphragm welded tightly to the head's body (differential pressure transmitters have two separated diaphragms for the inputs: "+" and "-"). The measuring head is mounted in the housing and secured with two screws.

In the heads to measure differential pressure and absolute pressure the tight bushings are applied. For overpressure measurements at a pressure range head to 7MPa, bushings are used with the opening from which a tube connecting the rear side of the measuring diaphragm to the atmosphere is pulled out; there are cylindrical flameproof joints used additionally in this case and in some versions of pressure difference heads. The transmitters with the head versions described above have category 1/2G, 1/2D.

In the versions pressure transmitters 3400-... and differential pressure transmitters 3500-... of category 2G and 2D (measured in zone 1 or 21) all pressure heads are allowed, including those without additional flame-proof joints.

Enclosures of transmitters are made of die-cast aluminium alloy or stainless steel. Enclosure consists of a body and two screwed covers (display and electrical connection). The cable line is introduced the enclosure flameproof cable gland with thread M20x1,5 or 1/2NPT depending on the version of the housing body. In the non-used opening the explosion-proof plug (cap) prod. DWYER INSTRUMENTS INC. is mounted.

The transmitters may be fitted with diaphragm seals, which enable them to be used in a variety of conditions such as thick or highly reactive media, high and low temperatures. Elements of the diaphragm seals can be coated with teflon.

In the name of transmitters types the letters "XX" stand for the symbol of enclosure type used.

- AL - in case of an aluminum enclosure
- AS - in case of an stainless steel enclosure

In the name of transmitters types the letters "YY" stand for the values based on the base range.



[13]  
[14]

**SCHEDULE**  
EU-type Examination Certificate  
**KDB 17ATEX0056X issue 0**



**Technical parameters:**

Range of the measured pressure:

-14,5 psi - 12325 psi	(type 3400-xx-01...43)
-1 bar - 850 bar	(type 3400-xx-60...94)
-9 in w.c. - 750 in w.c.	(type 3500-xx-02...15 and 50, 52)
-5 psi - 1000 psi	(type 3500-xx-20...40 and 60, 62)
-0,5 bar - 70 bar	(type 3500-xx-70...78 and 83, 84)
-7 mbar - 70 mbar	(type 3500-xx-80...82)

Output signal:

4-20mA in a two-wire system + HART

Supply voltage:

13,5V - 55V

Ingress protection:

IP66 / IP67

Ambient temperature:

-40°C - 45°C (-40°F - 113°F)	for T6, T85°C (T185°F)
-40°C - 75°C (-40°F - 167°F)	for T5, T100°C (T212°F)

**[16] Test report:**

„Sprawozdanie z oceny ATEX” KDB Nr 17.072

**[17] Special conditions for safe use:**

- Some of the flameproof joints dimensions are different than specified in standard EN 60079-1. The relevant information for the user is included in the manual;
- In areas where there is a risk of dust explosion, transmitters in aluminium alloy casing covered with lacquer and transmitters with plastic rating plates or with diaphragm seals covered by teflon should be installed in a way to prevent electrostatic charging according to the operation manual.

**[18] Essential health and safety requirements:**

Met by compliance with standards listed below:

- EN 60079-0:2012 + A11:2013 (PN-EN 60079-0:2013-03 + A11:2014-03);
- EN 60079-1:2014 (PN-EN 60079-1:2014-12);
- EN 60079-11:2012 (PN-EN 60079-11:2012);
- EN 60079-26:2015 (PN-EN 60079-26:2015-04);
- EN 60079-31:2014 (PN-EN 60079-31:2014-10);

**Document's history:**

- EU-Type Examination Certificate KDB 17ATEX0056X issue 0, **this document.**

